

0/c

From

The Principal
Post Graduate Govt. College for Girls,
Sec-42, Chandigarh

To

Memo No. PGGCG-42/U.G.C. Project/ 7096⁸⁰¹⁵ 2014 Dated: 17/10/14

Subject:- Quotaion for purchase of Gel Documentation System

Sealed quotations are invited for the purchase of Gel Documentation System as per specification given below. The quotations duly sealed in an envelope and marked "Quotations for the purchase of Gel Documentation System" should reach the office of the undersigned on or before 10/11/2014 (3.00 PM.)

The bids should be submitted in two parts; Technical Bid and Financial Bid in separate sealed envelopes. The two sealed envelopes then would be contained in a bigger sealed Envelope.

Technical Bid Envelope: The Technical Bid envelope would contain Draft for Rs. 10000/- drawn in favor of the Principal, P.G. Govt. College for Girls, Sec-42, Chandigarh, payable at Chandigarh and complete specifications duly signed by the firm/tenderer. The Technical Bid would be rejected if the rates are mentioned in Technical Bid envelope.

Financial Bid Envelope: The rates would be mentioned only in the Financial Bids. The financial bid of only those firms, who qualify on the basis of Technical Bids, would only be opened.


DETAILS OF SPECIFICATION

GEL DOCUMENTATION SYSTEM

- The Gel Doc. EZ Gel Imaging system includes stain free tray, camera image resolution >4 megapixels, cables and detector should be CCD with compatible laptop.
- The system should have stain free technology.
- The system should have only one emission filter to reduce user-introduced errors.
- It may be used for detection methods like Ethidium bromide, SYBR® Green, Oriole, CY3, green fluorescent protein, silver stain, Coomassie Brilliant Blue etc.
- There should be reproducibility in results and the quality of images should be accepted for publications.
- Sophisticated software with auto image capture, auto analysis.
- SYSTEM should allow 100% repeatability of the workflow by any user and ensures optimized image data and analysis from a gel in a single uninterrupted, fast, and completely reproducible workflow.
- Power supply – 220/240 VAC, 50/60 Hz
- Operating Temperature - 10-28°C
- Operating Humidity - < 70% non-condensing
- Safety Compliance: EN61010
- Warranty – 3 year against any manufacturing defect.

Note:-

4. Please mention the complete specification, design/model no. of item to be supplied by you in Technical Bid envelop duly signed by the authorized signatory.
5. Please quote rate of item in Financial Bid.
6. EMD for Rs.10,000/- (Rs. Ten Thousand only) in the shape of Demand Draft drawn in favour of the Principal P.G. Govt. College for Girls Sec-42, Chandigarh payable at Chandigarh to be enclosed in Technical Bid Envelop.
4. Quotations should be complete in all respects
5. Rates of VAT, Sales Tax etc. if applicable should also be mentioned.
6. The rates should be F.O.R. College.
7. Incomplete quotations will be rejected.
8. Quotations should be sealed properly.
9. The performance /Bank Guarantee may also be submitted.
10. Details are also available on college website: www.gcg42.ac.in and college notice board.


Principal
PG Govt. College for Girls,
Sector-42, Chandigarh